

Warm-up Procedure

Step	Tube Voltage (kV)	Current-Time Product (mAs)	Number of Times	Pause (seconds)
1	60	10	2	40
2	70	10	2	40
3	80	10	2	40
4	90	10	2	40
5	100	10	2	40
6	110	10	2	40
7	115	10	2	40
8	120	10	2	40
9	125	10	2	40
10*	130	10	2	40
11*	133	10	2	40